

09-16-10 11:40

* Comparison of the detector findings with data derived from different imaging techniques during biopsy & the results of pathological studies & with the growing data of the electrical disturbances of different disorders (e.g. loss of the glomerular negative charges in Nephrotic Syndrome & recently the electrical disturbances in some hepatic disorders) will help to verify the significance of the detector & feed its computerized analysis units.

B-Manufacture.

- * Supply the sensor as a disposable sterile single use product.
- * The cable & the monitor can be manufactured as permanent devices. Alternatively a small electric lamp or ring can be added to the base of the sensor to be used for detection of physical & electrical resistance.

C-Utilization .

- * The sensors will be consumed with the rate of biopsies taken .
- * The monitors ,cables & data analysis units will be consumed with a number parallel to the number of the medical units performing biopsy .

HL
11-3-09